

KAPA SYBR® FAST qPCR Kits

Instrument Compatibility

| | | | ABI Prism® 7000 | ABI Prism® 7700 | ABI Prism® 7900 HT | ABI 5700 | ABI 7300 | ABI 7500 | ABI StepOne™/StepOnePlus™ | ABI ViiA™ 7 | Bio-Rad iCycler iQ®/iQ®5 | Bio-Rad iCycler MyiQ™ | Bio-Rad CFX96™/CFX384™ | Bio-Rad Opticon™ | Bio-Rad Chromo4™ | Cepheid Smart Cycler® | Corbett Rotor-Gene™ | Eppendorf Mastercycler® ep realplex | Helix Pixo™ | Qiagen Rotor-Gene™ Q | Roche LightCycler® 480/1536 | Stratagene Mx3000P™ | Stratagene Mx3005P™ | Stratagene Mx4000™ | |
|--|--------|-----------------------------|-----------------|-----------------|--------------------|----------|----------|----------|---------------------------|-------------|--------------------------|-----------------------|------------------------|------------------|------------------|-----------------------|---------------------|-------------------------------------|-------------|----------------------|-----------------------------|---------------------|---------------------|--------------------|--|
| KAPA SYBR® FAST qPCR (2x) Universal Kit: Includes separate tubes of ROX passive reference dye (high and low) | KK4600 | 1 x 1 ml (100, 20 µl rxns) | | | | | | | | | | | | | | | | | | | | | | | |
| | KK4601 | 1 x 5 ml (500, 20 µl rxns) | ○ | ○ | ○ | ○ | ○ | ● | ○ | ○ | | | ● | ● | ● | ● | ● | ● | ● | ● | ○ | ● | ● | ● | |
| | KK4602 | 2 x 5 ml (1000, 20 µl rxns) | | | | | | | | | | | | | | | | | | | | | | | |
| KAPA SYBR® FAST qPCR (2x) ABI Prism® Kit: ROX passive reference dye included in 2x Master Mix | KK4603 | 1 x 1 ml (100, 20 µl rxns) | | | | | | | | | | | | | | | | | | | | | | | |
| | KK4604 | 1 x 5 ml (500, 20 µl rxns) | ● | ● | ● | ● | ● | | ● | ● | | | | | | | | | | | | | | | |
| | KK4605 | 2 x 5 ml (1000, 20 µl rxns) | | | | | | | | | | | | | | | | | | | | | | | |
| KAPA SYBR® FAST qPCR (2x) Bio-Rad iCycler™: Fluorescein reference dye included in 2x Master Mix | KK4606 | 1 x 1 ml (100, 20 µl rxns) | | | | | | | | | | | | | | | | | | | | | | | |
| | KK4607 | 1 x 5 ml (500, 20 µl rxns) | | | | | | | | | ● | ● | | | | | | | | | | | | | |
| | KK4608 | 2 x 5 ml (1000, 20 µl rxns) | | | | | | | | | | | | | | | | | | | | | | | |
| KAPA SYBR® FAST qPCR (2x) Kit optimized for the LightCycler® 480: Contains no reference dye | KK4609 | 1 x 1 ml (100, 20 µl rxns) | | | | | | | | | | | | | | | | | | | | | | | |
| | KK4610 | 1 x 5 ml (500, 20 µl rxns) | | | | | | | | | | | | | | | | | | | | ● | | | |
| | KK4611 | 2 x 5 ml (1000, 20 µl rxns) | | | | | | | | | | | | | | | | | | | | | | | |

● Indicates preferred kit for this cycler

○ Indicates that kit can be used but may require the addition of ROX reference dye